



Jim Doyle, Governor
Mary P. Burke, Secretary

March 17, 2004

T.H.E. MEDICAL
JOE BAILEY
526 BRYNE DRIVE
BARRIE, ONTARIO L4N9P6
CANADA

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: T.H.E. MEDICAL
Product Name: LIBERTI SIDE ENTRY TUB
Model Number(s): TBL2000 and OASIS Side Entry Tub
Product File No: 20040089

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters Comm 82 through 84, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of March 2009.

This approval is contingent upon compliance with the following stipulation(s):

- The Mixing Panel serving this product must be mounted at a height that will place the critical level of the vacuum breakers serving the hand held shower and disinfectant injector at a elevation of at least 6 inches above the following:
 1. Highest use of the hand held shower and
 2. Point of injection of the disinfectant.

This approval letter shall be incorporated with your previously approved plans and/or specifications approved under product file number 20030350.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Michael J. Beckwith, CIPE
Plumbing Product Reviewer
phone: 608-266-6742
fax: 608-267-9566
e-mail: mike.beckwith@wi.gov